

ADVANCED BIOMEDICAL WASTE TREATMENT SOLUTIONS



France

www.tesalys.fr



STERIPLUS™ 20

integrated system with preliminary shredding and steam sterilization of your biohazardous waste





U	ı	I	L	I	I	I	E	S	
W	ı	Δ	т	F	1	2			

REQUIREMENTS	
CONSUMPTION	

Softened water 4-8° TH 5 liters / cycle

ELECTRICITY

REQUIREMENTS CONSUMPTION

380-400 V III N. 50Hz/60Hz. 10 kW 1 kW/h / cycle

DRAIN

32 mm ø internal (heat-resistant)

STEAM/COMPRESSED AIR

Produced by the integrated steam generator and air compressor; no external connection required.

(1) 30-35min for test loads as per NF X30-503-1 standard.

(2) Cycle times may vary based on the type of materials, waste density, humidity, quantity of liquids, etc. For loads with high liquid content, cycle time can be longer. Up to 50min depending on the waste density.

(3) Based on an average cycle time of 45min. (40 min. of cycle + 5 min. unloading/loading time).

(4) Quantity of liquids allowed might vary depending on application (urine bags, blood bags, food testing bags, ...). All the data given in this catalogue are non-binding and given as a guide only. Treatment capacities, cycle times and consumptions are given as mere examples and may vary depending on the type of waste, installation conditions, operator's availability and skills.

Consumption is calculated on the basis of a standard cycle of the machine. For consumption with cooling of the drain and / or the Virus + kit, please consult us.

LOADING CHAMBER CAPACITY

LOADING CHAMBER USEFUL DIMENSIONS (W x D x H)

320 x 247 x 273 mm

201

LOADING CAPACITY (CONTAINERS)

TOTAL CYCLE TIME (NF X30-503-1 STANDARD) (1)

1 x TESABOX 20 30 to 35 min

TOTAL CYCLE TIME IN WORKING

CONDITIONS (2)

30 to 50 min

Based on an average waste density of 0,08 to 0,12kg/L for Healthcare Waste

LOADING CAPACITY (KG/CYCLE)

1,6 to 2,4 kg/cycle

Based on an average waste density of 0,12 to 0,2kg/L for Lab/Humid/Liquid Waste

2,4 to 4 kg/cycle

Based on an average waste density of 0,08 to 0,12kg/L for Healthcare Waste

2 to 3 kg/h

Based on an average waste density of 0,12 to 0,2kg/L for Lab/Humid/Liquid Waste

3 to 5 kg/h

MAX QUANTITY OF LIQUID **RECOMMENDED PER CYCLE (4)**

TREATMENT CAPACITY (KG /H) (3)

6 L

EXTERNAL DIMENSIONS (W x D x H)

1300 x 895 x 1210 mm

NET WEIGHT

530 ka

LOADING CHAMBER

Stainless steel with PTFE coating

TREATMENT CHAMBER

AISI 304L stainless steel

AISI 304L stainless steel (chamber) & Highly

STEAM GENERATOR

corrosion-resistant Hastelloy®

WATER SOFTENER

Optional (external)

WATER BOOSTER COMPRESSOR

PUMP AND AIR

Structure: AISI 304L stainless steel & Blades:

WASTE SHREDDER

High-strength steel Stainless steel/PTFE flexible hosing

PIPING CHASSIS

Painted steel

Composite material/Painted steel

CYCLE TRACEABILITY Through printout

Built-in

USB PORT

Built-in

STERIPLUS™ A COMPLETE RANGE



STERIPLUS™ 20 2/5 kg/h



STERIPLUS™ 40 4/10 kg/h



STERIPLUS™ 80 8/20 kg/h



TESABAG Autoclavable biohazardous waste bags



TESABOX Cardboard containers for biohazardous waste