

ADVANCED BIOMEDICAL WASTE TREATMENT SOLUTIONS



France

www.tesalys.fr



STERIPLUS™ 40

integrated system with preliminary shredding and steam sterilization of your biohazardous waste





JTILITIES	
WATER	
REQUIREME CONSUMPT	Softened water 4-8° TH 10 liters / cycle
ELECTRICITY	
REQUIREME CONSUMPT	380-400 V III N, 50Hz/60Hz, 15 kW 1.5 kW/h / cycle
DRAIN	32 mm ø internal (heat-resistant)

STEAM/COMPRESSED AIR

Produced by the integrated steam generator and air compressor; no external connection required.

- (1) 30-35min for test loads as per NF X30-503-1 standard.
- (2) Cycle times may vary based on the type of materials, waste density, humidity, quantity of liquids, etc. For loads with high liquid content, cycle time can be longer. Up to 50min depending on the waste density.
- (3) Based on an average cycle time of 45min. (40 min. of cycle + 5 min. unloading/loading time).
- (4) Quantity of liquids allowed might vary depending on application (urine bags, blood bags, food testing bags, ...). All the data given in this catalogue are non-binding and given as a guide only. Treatment capacities, cycle times and consumptions are given as mere examples and may vary depending on the type of waste, installation conditions, operator's availability and skills.

Consumption is calculated on the basis of a standard cycle of the machine.

For consumption with cooling of the drain and / or the Virus + kit, please consult us.

LOADING CHAMBER CAPACITY	40 L
LOADING CHAMBER USEFUL DIMENSIONS (W x D x H)	320 x 247 x 520 mm
LOADING CAPACITY (CONTAINERS)	1 x TESABOX 40 or 2 x TESABOX 20
TOTAL CYCLE TIME (NF X30-503-1 Standard) (1)	30 to 35 min
TOTAL CYCLE TIME IN WORKING CONDITIONS (2)	30 to 50 min
	Based on an average waste density of 0,08 to 0,12kg/L for Healthcare Waste
LOADING GARAGITY (VG (OVG) E)	3,2 to 4,8 kg/cycle
LOADING CAPACITY (KG/CYCLE)	Based on an average waste density of 0,12 to 0,2kg/L for Lab/Humid/Liquid Waste
	4,8 to 8 kg/cycle
	Based on an average waste density of 0,08 to 0,12kg/L for Healthcare Waste
TOTATALENT OADAOITY (VO. (II) (O)	4 to 6 kg/h
TREATMENT CAPACITY (KG /H) (3)	Based on an average waste density of 0,12 to 0,2kg/L for Lab/Humid/Liquid Waste
	6 to 10 kg/h
MAX QUANTITY OF LIQUID RECOMMENDED PER CYCLE (4)	12 L

1300 x 895 x 1450 mm

AISI 304L stainless steel

Built-in

Built-in

High-strength steel

Painted steel

Built-in

Built-in

Stainless steel with PTFE coating

corrosion-resistant Hastelloy®

AISI 304L stainless steel (chamber) & Highly

Structure: AISI 304L stainless steel & Blades:

Stainless steel/PTFE flexible hosing

Composite material/Painted steel

STERIPLUS™ A COMPLETE RANGE



STERIPLUS™ 20 2/5 kg/h



STERIPLUS™ 40 4/10 kg/h



STERIPLUS™ 80 8/20 kg/h



EXTERNAL DIMENSIONS (W x D x H)

WATER BOOSTER PUMP AND AIR

CYCLE TRACEABILITY Through printout

NET WEIGHT LOADING CHAMBER

TREATMENT CHAMBER

STEAM GENERATOR

WATER SOFTENER

WASTE SHREDDER

PIPING

CHASSIS

BODYWORK

USB PORT

TESABAG Autoclavable biohazardous waste bags



TESABOX
Cardboard containers
for biohazardous waste